



: 2538-2664/0203

Fax: 2538-6442

MADHYAMGRAM MUNICIPALITY

P.O.: Madhyamgram, Dist.: North 24 Parganas

Kolkata – 700 129

Ref No: MM/E.O./HOSP/INSTRUMENT/NIQ-092/25-26

Date: 22/07/2025

NOTICE INVITING QUOTATION

Sealed Quotation are therefore invited from Manufacturer/Authorized agent, the rate should be quoted as per pack size as mentioned in the list and should **inclusive GST**. The Quotation should be submitted in Quotation form duly sealed covered super scribing “ **Quotation of rate for supply of INSTRUMENT Items**” in ‘Netaji Subhas Chandra Bose Specialised Hospital & Research Centre’, **addressed to the Chairman, Madhyamgram Municipality**, so as to reach this office within 12.30 Hrs. on 31/07/2025. The Quotation received after due date & time will not be considered. The Quotation will be opened on 31/07/2025 at 1.30 Hrs. and the Quotation are requested to remain present or send their suitable authorized persons to be present at the time of opening and offer clarification required if any.

Terms & Condition:-

1. Quotation forms will be available at the store of the NSCB SHOSPITAL & RC under Madhyamgram Municipality .
2. The last date of collecting Quotation From on – 31/07/2025 within 12.30hrs.
3. Supply against order to the successful Quotation is to be made as when as per requirements.
4. Supplier should remain bound to follow all other terms & condition as will be specified in the supply order.
5. Supplier must be submitted up to date Xerox copy of their up to date trade license, GST Registration, Pan Card, P. Tax Certificate & Income Tax.
6. Authority reserve the rights of accept or reject or every tender without assigning reason whatsoever.
7. First priority first brand name.
8. Quotation dropping address at Finance Officer's Room (Madhyamgram Municipality)
9. The above Quotation is also available on our website, www.madhyamgrammunicipality.org.

Executive Officer
Madhyamgram Municipality

Executive Officer
Madhyamgram Municipality
North 24 Parganas



: 2538-2664/0203

Fax: 2538-6442

MADHYAMGRAM MUNICIPALITY

P.O.: Madhyamgram, Dist.: North 24 Parganas

Kolkata – 700 129

Ref No: MM/E.O./HOSP/MONITOR/NIQ-092/25-26

Date: 22/07/2025

ANNEXURE-I

SL NO	Name of the Articles	Pack size
1	<u>MULTIPARA PATIENT MONITOR</u>	1 NO.

FOLLOWING SPECIFICATION & FEATURES ARE REQUIRED FOR MULTIPARA PATIENT MONITOR FOR GYNO WARD.

TECHNICAL SPECIFICATION OF MULTIPARA PATIENT MONITOR

POWER SUPPLY – 100-240 V, 50/60/Hz, adopts 12.1” color TFT LCD DISPLAYING REAL –TIME DATE AND WAVE FORM. SYNCHRONOUSLY DISPLAY EIGHT - CHANNEL WAVEFORM AND FULL MONITORING PARAMETERS EQUIPPED WITH AN OPTIONAL 48MM THERMAL RECORDER.

.REPLACEABLE INTERNAL BATTERY BRINGS LOT OF CONVENIENT FOR PATIENT MOVING.

THE REQUIRED MONITOR DEVICE SUCH PARAMETERS AS ECG, RESP, NIBP, SPO2, & DUAL – CHANNEL TEMP., ETC.

FEATURES :

- 17) 12.1 INCH TFT COLOR LCD , MULTI LANGUAGE INTERFACE.
- 18) FANLESS DESIGN , QUIET, ENERGY- SAVING AND CLEAN, WHICH REDUCES THE POSSIBILITY OF CROSS- INFECTION.
- 19) ALL- ROUND MONITOR FOR ADULT , PEDIATRIC AND NEONATE.
- 20) WITH STANDERD INTERFACE , OXYGEN GRAPH, TREND GRAPH, BIG CHARACTER INTERFACE AND VIEW BED, CONVENIENT TO OBSERVE.
- 21) FINISH ALL OPERATION BY KEYS AND KNOBS.
- 22) MAXIMUM 8- CHANNEL WAVEFORM SYNCHRONOUS DISPLAY.
- 23) DISPLAY 7-LEAD ECG WAVEFORM ON ONESCREEN, CASCADE ECG WAVEFORM DISPLAY.
- 24) ADOPT DIGITAL SPO2 TECHNOLOGY, ANTIMOTION AND ANTI- AMBIENT LIGHT INTERFERENCE, AND MEASURMENT CAN BE PERFORMED UNDER THE CIRCUMTANCE OF WEAK FILLING.
- 25) HEART RATE VARIABILITY (HRV) ANALYSIS FUNCTION.
- 26) NIBP MEASURMENT MODE: MANUAL/AUTO/STAT, STORAGE FOR 4800-GROUP NIBP DATA.
- 27) REVIEW FOR 71 ALARM EVENTS OF ALL PARAMETERS AND 60 ARRYTHMIA ALARM EVENTS.
- 28) DRUG CONNECTION CALCULATION AND TITRATION TABLE FUNCTIONS.
- 29) ONE-TOUCH PRINTING OF TREND GRAPH.
- 30) CONNECT TO CENTRAL MONITORING SYSTEM BY 3G, WIFI OR WIRED MODE.
- 31) AC/DC, BUILT- IN RECHARGEABLE LITHIUM BATTERY ACHIEVE UNINTERRUPTED MONITORING.
- 32) ANTI- HIGH FREQUENCY SURGICAL UNIT, DEFIBRILLATION – PROOF (SPECIAL LEAD ARE NECESSARY).

PERFORMANCE

ECG

LEAD MODE: 3LEAD OR 5LEAD

LEAD SELECTION: I,II,III, aVR,aVL,aVF,V

Wave form : 5 lead, dual – channel, 3 lead, single channel

Gain : 2.5mm/mV, 5.0mm/mV, 25mm/s, 50mm/s

HR: measurement and alarm range : 15 to 350bpm

Accuracy: (+,-) 1% or (+,-)1bpm, whichever is greater

Alarm accuracy : (+,-)2 bpm

Resolution : 1 bpm

St- segment monitoring:

Measurement & alarm range : -2.0mVto +2.0mV

Accuracy : -0.8mV to + 0.8mV(+,-)0.04mV or +,-10% , whichever is greater

Arrhythmia analysis : ASYSTOLE, VFIB/VTAC, COUPLET, BIGEMINY, TRIGEMINY, RONT , VT>2,PVC, TACHY, BRADY, MISSED BEATS, PNP, PNC

Thanking you,



Executive Officer

Madhyamgram Municipality

Executive Officer

Madhyamgram Municipality

North 24 Parganas